



Product Service

EC-CERTIFICATE

Full Quality Assurance System

(Annex I, section 3 of the Directive 93/42/EEC on Medical Devices)

No. G1 09 04 44849 007

Manufacturer: Philips Medical Systems
3000 Minuteman Road
Andover MA 01810-1090
USA

EC-Representative: Philips Medizin Systeme
Böblingen GmbH
Hewlett-Packard Strasse 2
71043 Böblingen
GERMANY

Product Category(ies): Single and Multi-Parameter Patient Monitors (Transport, Point-of-Care, Central Unit, Home-Use);
Cardiographs; ICG Systems; ECG Systems; EEG Systems;
Telemetry Systems; Defibrillators and related Accessories (including Cables, Electrodes, Recorders) for ECG, Cardiac Output, ICG, EEG, CPR Feedback;
Oxygen Saturation, Respiration and Gas Monitoring;
Weight, Blood Pressure (Invasive and Non-Invasive),
Temperature, Physiological Pressure Kits,
Defibrillation (including Paddles) and Clinical Information Systems;
Transnasal Sheath Kits

The Certifier Body of TÜV SÜD Product Service GmbH declares that the aforementioned manufacturer has implemented a quality assurance system for design, manufacture and final inspection of the respective products / product categories according to Annex I section 3 of the Directive 93/42/EEC on Medical Devices. This quality assurance system conforms to the provisions of the Directive and is subject to periodical surveillance. For marketing of class II products an additional Annex I certificate is mandatory. See also notes on rear!

Report No. DM00017

Valid until 2014-08-05

Date 2009-08-08

Hans-Herbert Junker



TÜV SÜD Product Service GmbH is Notified Body according to Council Directive 93/42/EEC concerning medical devices with identification no. 0123

Page 1 of 2